HEMOSIL®

LUPUS ANTICOAGULANT (LA) ASSAY PANEL



werfen

Does your LA testing meet the standard?

Guidelines for Lupus Anticoagulant (LA) testing call for two types of tests, Dilute Russell Viper Venom Test (dRVVT) and an APTT-based assay with a silica activator.* Together, these tests optimize sensitivity and specificity in LA. Only IL offers both tests on a single, fully automated panel.

HemosIL LA assay panel includes Silica Clotting Time (SCT), the only silica-based APTT assay that, when combined with dRVVT, offers increased sensitivity1 and results on one system. Meet the new standard, with HemosIL.

The only automated LA panel with both SCT and dRVVT

- Features APTT-based assay with silica activator as recommended by ISTH SSC 2009 LA testing guidelines
- SCT and dRVVT testing, together, offer enhanced detection:
  - More likely to differentiate LA from anti-FVIII inhibitors than either test alone2,3
  - Highest sensitivity in the detection of LA in patients who meet the clinical criteria for antiphospholipid syndrome (APS)1
- Combined with ACL AcuStar® Hemostasis Testing System, covers the maximum spectrum of antiphospholipid (aPL) antibody detection

Efficient and compliant

- Easy to use, fully automated on IL Hemostasis Testing Systems
- Screen and Confirm available for both methodologies
- Fully compliant with guidelines for LA detection
- Results expressed as Normalized Ratio (NR)



*Recommended by ISTH SSC 2009 LA testing guidelines

Excellent stability and performance on ACL TOP Systems

HemosIL LA panel analytical performance

	Silica Clotting Time	dRVVT Screen and Confirm	LA Negative Control	LA Positive Control
Stability at 2–8° C	20 Days	15 Days	24 Hours	24 Hours
Onboard Stability†	5 Days	3 Days	24 Hours	24 Hours
Expected Values	Normalized SCT Ratio	Normalized dRVVT Ratio	Normalized SCT Ratio	Normalized dRVVT Ratio
ACL TOP 700	0.84–1.16	0.92–1.11	<1.16	≥1.40
ACL ELITE®/ ELITE PRO	0.75–1.20	0.91–1.14	<1.20	≥1.40

HemosIL LA Negative Control: ideal for mixing studies

- LA Controls prepared from human samples screened +/- according to ISTH guidelines4
- In-house study demonstrated similar clotting times between LA Negative Controls and Platelet Poor Normal Pooled Plasma (PPNPP) measured daily over a 7-day period using HemosIL LA and APTT assays
- LA Negative Control can replace Pooled Normal Plasma in LA mixing studies

	Silica Clotting Time	dRVVT Screen and Confirm	SynthASil	APTT-SP
PPNPP Mean Clot Time (sec)	32.8 (Screen) 35.7 (Confirm)	32.8 (Screen) 29.0 (Confirm)	32.0	30.1
LA Negative Control Mean Clot Time (sec)	33.3 (Screen) 36.9 (Confirm)	32.8 (Screen) 30.6 (Confirm)	30.1	29.1

†Onboard ACL TOP Family Systems

Screening for LA

LA diagnosis is a complex procedure


ISTH SSC 2009 LA testing guidelines4

- Recommends two phospholipid-dependent clotting assays based on different methodologies to demonstrate phospholipid dependence
- Only HemosIL includes both assays in one panel, and more

LA Testing Recommendations4	HemosIL Solution
dRVVT Assay Specific for LA in patients at high risk for thrombosis	HemosIL dRVVT Screen/Confirm
Silica-based APTT Assay Low in phospholipid concentration, selected for its sensitivity	HemosIL Silica Clotting Time
Established +/- Controls Must be used	HemosIL LA Positive/Negative Controls
Mixing Test Pooled Normal Plasma (fresh or lyophilized)	LA Negative Control

Laboratory criteria

- Prolongation of phospholipid-dependent assays
- Exclusion of interfering factors (factor deficiencies, factor inhibitors, heparin and oral anticoagulants)
- Determination of phospholipid dependence




One standard. One solution.

Enhanced LA testing sensitivity

- dRVVT and SCT enhance the identification of LA with different sensitivities to LA antibodies
- dRVVT and SCT combined broadens the detection capability, resulting in the highest sensitivity


Sensitivity and specificity in LA testing1



	SCT	APTT-LA	DRVVT	DRVVT/APTT-LA	DRVVT/SCT
Sensitivity	53.3	31	31	48.8	57.7
Specificity	96.6	100	98.9	98.9	96.6

With APL testing, a complete APS diagnostic assessment

LA testing on the ACL TOP together with APL testing on ACL AcuStar meets ISTH guidelines for a complete APS diagnosis.



BETTER TOGETHER. Meet the standard for LA testing with the only single, fully automated panel that offers both dRVVT and SCT. HemosIL.

** Both clinical and laboratory criteria must be met in order to diagnose APS. Refer to ISTH Guidelines on APS Diagnosis 2006.

HemosIL Lupus Anticoagulant Assay Panel

Product	Part Number	Kit Configuration
Silica Clotting Time	0020004800	SCT Screen 3 x 5 mL (liq) SCT Confirm 3 x 5 mL (liq) SCT CaCl2 3 x 10 mL (liq)
dRVVT Screen	0020301500	10 x 2 mL (lyo)
dRVVT Confirm	0020301600	10 x 2 mL (lyo)
LA Positive Control	0020012500	10 x 1 mL (lyo)
LA Negative Control	0020012600	10 x 1 mL (lyo)

References

1. Grypiotis P, Ruffati A, *et al*. Use of a new silica clotting time for diagnosing lupus anticoagulant in patients who meet the clinical criteria for antiphospholipid syndrome. *J Clin Lab Analysis* 2006;20:15–18.
2. Tripodi A, Mancuso ME, *et al*. Lupus anticoagulants and their relationship with the inhibitors against coagulation FVIII: considerations on the differentiation between the 2 circulating anticoagulants. *Clin Chem* 2005;10:1883–5.
3. Chantarangkul V, Tripodi A, *et al*. Silica Clotting Time (SCT) as a screening and confirmatory test for detection of the lupus anticoagulants. *Thromb Res*. 1992;67:355–65.
4. Pengo V, Tripodi A, Reber G, Rand H, Ortel TL, Galli M and DeGroot PG. Update of the guidelines for lupus anticoagulant detection. *Thromb Haemost*. 2009 Oct 7(10):1737–40.
5. Miyakis S, Lockshin MD, *et al*. International consensus statement on an update of the classification criteria for definite antiphospholipid syndrome (APS). *J Thromb Haemost*. 2006;4:296-306.

Additional Literature

- Arnout J. Antiphospholipid syndrome: diagnostic aspects of lupus anticoagulants. *Thromb Haemost*. 2001;86:83–91.
- Brandt JT, Triplett DA, *et al*. Criteria for the diagnosis of lupus anticoagulants: an update. *Thromb Haemost*. 1995;74:1185–90.
- Chantarangkul V, Tripodi A, *et al*. Laboratory diagnosis of lupus anticoagulants. *Thromb Haemost*. 2002;87:854–8.
- Dragoni F, Minotti C, *et al*. As compared to kaolin clotting time, silica clotting time is a specific and sensitive method for detecting lupus anticoagulant. *Thromb Res*. 2001;101:45–51.
- Galli M, Dlott J, *et al*. Lupus anticoagulants and thrombosis: clinical association of different coagulation and immunological tests. *Thromb Haemost*. 2000;84:1012–6.
- Galli M. Which antiphospholipid antibodies should be measured in the antiphospholipid syndrome? *Haemostasis*. 2000;30 Suppl 2:57–62.
- Horbach DA, Van Oort E, *et al*. Lupus anticoagulant is the strongest risk factor for both venous and arterial thrombosis in patients with systemic lupus erythematosus. Comparison between different assays for the detection of antiphospholipid antibodies. *Thromb Haemost*. 1996;76:916–24.
- International consensus statement on an update of the classification criteria for definite antiphospholipid syndrome (APS). *J Thromb Haemost*. 2006 Feb;4(2):295–306.
- Montaruli B, Vaccarino A, *et al*. Lupus anticoagulant: performance of a new, fully automated commercial screening and confirmation assay. *Clin Chem*. 2005;6:1031–3.
- Whal DG, Guillemin F, *et al*. Risk of venous thrombosis related to antiphospholipid antibodies in systemic lupus erythematosus. *Lupus*. 1997;6:467–73.
- Wilson WA, Gharavi AE, *et al*. International consensus statement on preliminary classification criteria for definite antiphospholipid syndrome: report of international workshop. *Arthritis Rheum*. 1999;42:1309–11.

Werfen
Corporate Headquarters
Plaza de Europa, 21–23
08908 L'Hospitalet de Llobregat
Barcelona, Spain
+34-93-4010101
werfen.com

Instrumentation
Laboratory Headquarters
180 Hartwell Road
Bedford, MA 01730 USA
+1-781-861-0710
www.ilww.com

Worldwide Locations

US, Canada, Latin America and South America

Brazil
São Paulo
+55-11-41543337
br.werfen.com

Canada
Richmond Hill, ON
+1-800-552-2025 x6115
www.ilus.com

Colombia
Bogotá
+57(1)-616-7513

Mexico
Col. Granada
+52-55-5262-1760
ildiagnostics.com

Uruguay
Montevideo
+5982-481-81-33

USA
Bedford, MA
+1-781-861-0710
www.ilus.com

Pacific

Australia
Artarmon, Sydney
+61-02-9098-0200

China
Shanghai
+86-21-66308671
Beijing
+86-10-59756055
cn.werfen.com

Hong Kong
Hong Kong
+852-2792-7773
cn.werfen.com

India
New Delhi
+91-490-29-550
in.ilwerfen.com

Japan
Minato-ku, Tokyo
+81-3-5419-1301
jp.werfen.com

Korea
Seoul
+82-2-5719246
kr.werfen.com

Thailand
Bangkok
+66-271-226-28/9

Europe, Middle East, Africa

Austria
Vienna
+43-1-250-46-000
at.werfen.com

Benelux
Belgium
+32-2-7252052
The Netherlands
+31-76-5480100
benelux.werfen.com

Czech Republic
Prague
+420-246-090-931
cz.werfen.com

France
Le Pré Saint Gervais
+33-182-30-86-00
fr.werfen.com

Germany
Munich
+49-89-909070
de.werfen.com

Hungary
Budapest +36-1-882-73-10
hu.werfen.com

Italy
Milan
+39-02-25221
it.werfen.com

Lithuania
Kaunas
+370-37-313157
lt.werfen.com

Poland
Warsaw
+48-22-336-18-00

Portugal
Carnaxide
+351-214247312
pt.werfen.com

Russia
Moscow
+7-499-124-45-59
ru.ilwerfen.com

Spain
Barcelona
+34-902-20-30-90
es.werfen.com

UK
England
+44-1925-810141
uk.werfen.com

For all other countries visit
international.werfen.com

The Instrumentation Laboratory logo, HemosIL, ACL, ACL AcuStar, ACL ELITE and ACL TOP are trademarks of Instrumentation Laboratory Company and/or one of its subsidiaries or parent companies and may be registered in the United States Patent and Trademark Office and in other jurisdictions. All other product names, company names, marks, logos and symbols are trademarks of their respective owners. ©2015 Instrumentation Laboratory. All rights reserved.

For more information, contact your local Werfen sales representative or distributor.